

New England Biolabs Certificate of Analysis

Product Name: Luna[®] SARS-CoV-2 RT-qPCR Multiplex Assay Kit
 Catalog Number: E3019S
 Packaging Lot Number: 10149316
 Expiration Date: 02/2023
 Storage Temperature: -20°C
 Specification Version: PS-E3019S/L v1.0

Luna [®] SARS-CoV-2 RT-qPCR Multiplex Assay Kit Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
S0631AVIAL	SARS-CoV-2 Primer/Probe Mix (N1/N2/RP)	10139910	Pass
N2117SVIAL	SARS-CoV-2 Positive Control (N gene)	10143730	Pass
M3019SVIAL	Luna [®] Probe One-Step RT-qPCR 4X Mix with UDG	10142626	Pass
B1502AVIAL	Nuclease-free Water	10145291	Pass

Assay Name/Specification	Lot # 10149316
Functional Testing (RT-qPCR, SARS-CoV-2) Luna [®] SARS-CoV-2 RT-qPCR Multiplex Assay Kit is functionally tested and compared to a control lot using 10 ng of human total RNA spiked with synthetic SARS-CoV-2 RNA in an RT-qPCR assay, resulting in a standard curve with a calculated qPCR efficiency of 90-110%, and a dynamic range of 5 orders of magnitude.	Pass
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in Luna [®] SARS-CoV-2 RT-qPCR Multiplex Assay Kit and meet the designated specifications.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Christie Vazquez

Christie Vazquez
Production Scientist
14 Jun 2022

Michael Tonello

Michael Tonello
Packaging Quality Control Inspector
14 Jun 2022