

New England Biolabs Certificate of Analysis


Product Name: *Monarch® PCR & DNA Cleanup Kit (5 µg)*
Catalog Number: *T1030L*
Packaging Lot Number: *10208654*
Expiration Date: *10/2025*
Storage Temperature: *25°C*
Specification Version: *PS-T1030S/L v2.0*

Monarch® PCR & DNA Cleanup Kit (5 µg) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T1034-2	Monarch® DNA Cleanup Columns (5 µg)	10208180	Pass
T1032-3	Monarch® DNA Wash Buffer	10208179	Pass
T1031-3	Monarch® DNA Cleanup Binding Buffer	10208174	Pass
T1018-2	Monarch® Collection Tubes (2 ml)	10208163	Pass
T1016-2	Monarch® DNA Elution Buffer	10208157	Pass

Assay Name/Specification	Lot # 10208654
Functional Testing (DNA Recovery and Purity) Twenty-four Monarch® DNA Cleanup Columns are tested with 1 µg of 1 Kb Ladder resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	Pass
Functional Testing (Fragment Purification and Restriction Digestion) A linear pLA3 fragment is added to 8 Monarch® DNA Cleanup Columns, eluted and digested with DraIII-HF® resulting in ≥90% conversion to two-fragment pattern observed by gel electrophoresis in ≥75% of the samples.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Nikki Samuelson
Production Scientist
14 Dec 2023



Julie Gaumont
Packaging Quality Control Inspector
09 Feb 2024