

New England Biolabs Certificate of Analysis


Product Name: *Monarch® HMW DNA Extraction Kit for Tissue*
Catalog Number: *T3060S*
Packaging Lot Number: *10231545*
Expiration Date: *03/2026*
Storage Temperature: *25°C*
Specification Version: *PS-T3060S/L v3.0*

Monarch® HMW DNA Extraction Kit for Tissue Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T3062-21	Monarch® Protein Separation Solution	10231333	Pass
T3061-21	Monarch® HMW gDNA Tissue Lysis Buffer	10231331	Pass
T3056-21	Monarch® gDNA Elution Buffer II	10231329	Pass
T3018-21	Monarch® RNase A	10230203	Pass
T3015-21	Monarch® gDNA Wash Buffer	10231310	Pass
T3005-21	Monarch® DNA Capture Beads (10)	10231290	Pass
T3004-21	Monarch® Bead Retainers	10231522	Pass
T3003-21	Monarch® 2 ml Tubes (10)	10119955	Pass
T3002-21	Monarch® Pestles (5)	10119948	Pass
T3001-21	Monarch® Pestle Tubes (5)	10119944	Pass
T2118-23	Monarch® Spin Collection Tubes	10231419	Pass
P8200GVIAL	Proteinase K, Molecular Biology Grade	10234388	Pass

Assay Name/Specification	Lot # 10231545
<p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in Monarch® HMW DNA Extraction Kit for Tissue and meet the designated specifications.</p>	Pass
<p>Functional Testing (Genomic DNA Yield and Integrity Analysis) (Tissue) Genomic DNA is purified from 8 samples of NEB 10-beta Competent E. coli using the High Molecular Weight DNA Extraction from Bacteria protocol. Each set of reagents is functionally tested to ensure successful isolation of gDNA from 1 x 10⁹ cells. Yield of genomic DNA is ≥6 µg in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples, and DIN values are ≥9 in ≥80% of the samples. In ≥ 80% of samples, ≥ 75% of the DNA is 50 kb in size or greater.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Darcie Spaulding
Production Scientist
02 Apr 2024



Julie Gaumont
Packaging Quality Control Inspector
02 Apr 2024