


New England Biolabs Certificate of Analysis

Product Name: Monarch Genomic DNA Purification Kit
Catalog Number: T3010S
Packaging Lot Number: 10059924
Expiration Date: 09/2021
Storage Temperature: 25°C
Specification Version: PS-T3010S/L v1.0

Assay Name/Specification	Lot # 10059924
<p>Functional Testing (Blood Cell Lysis Buffer, Monarch®) Genomic DNA is purified from 8 samples of whole pig blood using the Protocol for Extraction and Purification of Genomic DNA from Mammalian Whole Blood (non-nucleated), with $\geq 2 \mu\text{g}$ of gDNA being recovered in $\geq 80\%$ of the samples. OD 260/280 and 260/230 are ≥ 1.75 in $\geq 80\%$ of the samples.</p>	Pass
<p>Functional Testing (Cell Lysis Buffer, Monarch®) Genomic DNA is purified from 8 individual aliquots of HeLa cells (106 cells/aliquot) using the Protocol for Extraction and Purification of Genomic DNA from Cultured Cells, resulting in $\geq 3 \mu\text{g}$ of gDNA being recovered in $\geq 80\%$ of the samples. OD 260/280 and 260/230 are ≥ 1.75 in $\geq 80\%$ of the samples.</p>	Pass
<p>Functional Testing (DNA Recovery and Purity) Twenty-four Monarch® gDNA Purification Columns are tested with $5 \mu\text{g}$ of HindIII digested Lambda resulting in $\geq 75\%$ recovery in $\geq 80\%$ of the samples. OD 260/280 and 260/230 are ≥ 1.75 in $\geq 80\%$ of the samples.</p>	Pass
<p>Functional Testing (RNase A, Monarch®) A $10 \mu\text{l}$ reaction in NEBuffer 4 containing 40 ng of fluorescein labeled RNA transcript and RNase A is incubated at 37°C. After incubation for 5 minutes, complete disappearance of the RNA substrate occurs at $\leq 1.0 \mu\text{g/ml}$ of RNase A, as determined by gel electrophoresis using fluorescent detection.</p>	Pass
<p>Functional Testing (Tissue Lysis Buffer, Monarch®) Genomic DNA is purified from 8 NEB10-Beta samples treated using the Supplemental Protocol for Genomic DNA Purification from Gram-negative Bacteria, resulting in $\geq 5 \mu\text{g}$ of gDNA being recovered in $\geq 80\%$ of the samples. OD 260/280 and 260/230 are ≥ 1.75 in $\geq 80\%$ of the samples.</p>	Pass
<p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in Monarch® Genomic DNA Purification Kit and meet the designated specifications.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.



Ana Egana
Production Scientist
10 Jul 2020

Darcie Spaulding
Packaging Quality Control Inspector
10 Jul 2020