


New England Biolabs Certificate of Analysis

Product Name: Monarch RNA Cleanup Kit (10 µg)
 Catalog Number: T2030S
 Packaging Lot Number: 10059912
 Expiration Date: 09/2021
 Storage Temperature: 25°C
 Specification Version: PS-T2030S/L v1.0


Assay Name/Specification	Lot # 10059912
<p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in Monarch® RNA Cleanup Kit (10 µg) and meet the designated specifications.</p>	Pass
<p>Functional Testing (RNA Recovery and Purity) Twenty-four Monarch® RNA Cleanup Columns (T2037, 10 µg) are tested with 1 µg of rRNA resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.</p>	Pass
<p>Functional Testing (Enzyme Inactivation and Removal) Eight Monarch® RNA Cleanup Columns (T2037, 10 µg) are tested with 1 µg of rRNA and 2 units of DNase I. A 50 µl reaction in DNase I Reaction Buffer with 1 µg of pBR322 and 50% of the column eluate incubated for 10 minutes at 37°C results in >99.95% removal of DNase I activity in ≥75% of the samples.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Ana Egana
 Production Scientist
 31 Mar 2020



Darcie Spaulding
 Packaging Quality Control Inspector
 31 Mar 2020