

New England Biolabs Certificate of Analysis

Product Name: *Diluent B*
Catalog Number: *B8002S*
Concentration: *1 X Concentrate*
Lot Number: *10049738*
Expiration Date: *03/2022*
Storage Temperature: *-20°C*
Specification Version: *PS-B8002S v1.0*
Composition (1X): *300 mM NaCl , 10 mM Tris-HCl , 1 mM DTT , 0.1 mM EDTA , 50 % Glycerol , 500 µg/ml BSA, (pH 7.4 @ 25°C)*

Diluent B Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
B8002SVIAL	Diluent B	10036381	Pass


Assay Name/Specification	Lot # 10049738
Endonuclease Activity (Nicking) A 50 µl reaction in CutSmart® Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 10 µl of Diluent B incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart® Buffer containing 1 µg of PhiX174-HaeIII DNA and a minimum of 10 µl of Diluent B incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
pH (buffers/solutions) The pH of 1X Diluent B is between pH 7.3 and 7.5 at 25°C.	Pass
qPCR DNA Contamination (E. coli Genomic) A minimum of 1 µl of Diluent B is screened for the presence of E. coli genomic DNA using SYBR® Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome.	Pass
RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA	Pass

Assay Name/Specification	Lot # 10049738
and a minimum of 1 µl of Diluent B is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	

This product has been tested and shown to be in compliance with all specifications.



Derek Robinson
Production Scientist
06 Mar 2019



Michael Tonello
Packaging Quality Control Inspector
16 Aug 2019