

## New England Biolabs Certificate of Analysis

Product Name: NEBNext<sup>®</sup> RNA Depletion Core Reagent Set  
 Catalog Number: E7865X  
 Packaging Lot Number: 10198337  
 Expiration Date: 04/2025  
 Storage Temperature: -20°C  
 Specification Version: PS-E7865S/L/X v1.0

NEBNext <sup>®</sup> RNA Depletion Core Reagent Set Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7753-4VIAL	NEBNext <sup>®</sup> DNase I	10180721	Pass
E7752-4VIAL	NEBNext <sup>®</sup> Thermostable RNase H	10180722	Pass
E6317-4VIAL	Nuclease-free Water	10180725	Pass
E6315-4VIAL	DNase I Reaction Buffer	10180724	Pass
E6314-4VIAL	NEBNext <sup>®</sup> Probe Hybridization Buffer	10180726	Pass
E6312-4VIAL	RNase H Reaction Buffer	10180723	Pass

Assay Name/Specification	Lot # 10198337
<p><b>* Individual Product Component Note</b>            Standard Quality Control Tests are performed for each component included in NEBNext<sup>®</sup> RNA Depletion Core Reagent Set and meet the designated specifications.</p>	Pass
<p><b>Functional Testing (rRNA Depletion)</b>            NEBNext<sup>®</sup> RNA Depletion Core Reagent Set is functionally validated using commercially available human RNA (e.g. UHRR). After treatment with the kit, a high-throughput sequencing library is made and sequenced. Reads from this library are classified as being derived from ribosomal RNA based on sequence homology. This method produces libraries with &lt;10% rRNA.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.

*Christine Sumner*

---

Christine Sumner  
Production Scientist  
27 Jun 2023

*Josh Hersey*

---

Josh Hersey  
Packaging Quality Control Inspector  
27 Jun 2023