

New England Biolabs Certificate of Analysis


Product Name: dsRNA Ladder
Catalog Number: N0363S
Concentration: 500 µg/ml
Unit Definition: N/A
Packaging Lot Number: 10079153
Expiration Date: 07/2022
Storage Temperature: -20°C
Storage Conditions: 10 mM Tris-HCl, 1 mM EDTA, (pH 8.0 @ 25°C)
Specification Version: PS-N0363S v1.0

dsRNA Ladder Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
N0363SVIAL	dsRNA Ladder	10079154	Pass

Assay Name/Specification	Lot # 10079153
A260/A280 Assay The ratio of UV absorption of dsRNA Ladder at 260 and 280 nm is between 1.9 and 2.3.	Pass
Electrophoretic Pattern (Marker) The banding pattern of dsRNA Ladder on both a 6% polyacrylamide gel and a 2% agarose-TBE gel shows discrete, clearly identifiable bands at each fragment of the marker, when stained with Ethidium Bromide at a concentration of 0.5 µg/ml.	Pass
RNA Concentration (A260) The concentration of dsRNA Ladder is between 500 and 550 µg/ml as determined by UV absorption at 260 nm.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Bhairavi Jani
Production Scientist
03 Aug 2020



Josh Hersey
Packaging Quality Control Inspector
03 Aug 2020