

## New England Biolabs Certificate of Analysis

**Product Name:** PCR Marker  
**Catalog Number:** N3234S  
**Concentration:** 300 µg/ml  
**Unit Definition:** N/A  
**Packaging Lot Number:** 10230742  
**Expiration Date:** 02/2026  
**Storage Temperature:** -20°C  
**Storage Conditions:** 10 mM Tris-HCl (pH 8.0), 1 mM EDTA  
**Specification Version:** PS-N3234S/L v1.0

PCR Marker Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
N3234SVIAL	PCR Marker	10228586	Pass
B7025SVIAL	Gel Loading Dye, Purple (6X), no SDS	10221491	Pass

Assay Name/Specification	Lot # 10230742
<b>A260/A280 Assay</b> The ratio of UV absorption of PCR Marker at 260 and 280 nm is between 1.8 and 2.0.	Pass
<b>DNA Concentration (A260)</b> The concentration of PCR Marker is between 300 and 315 µg/ml as determined by UV absorption at 260 nm.	Pass
<b>Electrophoretic Pattern (Marker)</b> The banding pattern of PCR Marker on a 3% agarose gel shows discrete, clearly identifiable bands at each band of the marker, when stained with Ethidium Bromide at a concentration of 0.5 µg/ml.	Pass
<b>Non-Specific DNase Activity (DNA, 16 hour)</b> A 50 µl reaction in 1X NEBuffer 2 containing 1.5 µg of PCR Marker incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



Vanessa Mathieu-Sheltry  
Production Scientist  
26 Jan 2024



Michael Tonello  
Packaging Quality Control Inspector  
07 Mar 2024