

New England Biolabs Certificate of Analysis

Product Name: Bsp1286I
Catalog #: R0120S/L
Concentration: 5,000 units/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in 50 µl of reaction buffer.
Lot #: 0191605
Assay Date: 05/2016
Expiration Date: 5/2018
Storage Temp: -20°C
Storage Conditions:
Specification Version: PS-R0120S/L v1.0
Effective Date: 12 Apr 2013

Assay Name/Specification (minimum release criteria)	Lot #0191605
Exonuclease Activity (Radioactivity Release) - A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³ H] <i>E. coli</i> DNA and a minimum of 50 units of Bsp1286I incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) - After a 10-fold over-digestion of Lambda DNA with Bsp1286I, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Bsp1286I.	Pass
Non-Specific DNase Activity (16 Hour) - A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 5 Units of Bsp1286I incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

* The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.

M. W. Southworth

Authorized by
Maurice Southworth
12 Apr 2013

Penghua Zhang

Inspected by
Penghua Zhang
10 May 2016

