

New England Biolabs Certificate of Analysis

Product Name: Aval
Catalog Number: R0152M
Concentration: 50,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.
Packaging Lot Number: 10103905
Expiration Date: 03/2023
Storage Temperature: -20°C
Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA
Specification Version: PS-R0152T/M v1.0

Aval Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0152MVIAL	Aval	10103906	Pass
B6004SVIAL	rCutSmart™ Buffer	10103711	Pass

Assay Name/Specification	Lot # 10103905
<p>Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 100 Units of Aval incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.</p>	Pass
<p>Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 100 units of Aval incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.</p>	Pass
<p>Blue-White Screening (Terminal Integrity) A sample of pUC19 vector linearized with a 10-fold excess of Aval, religated and transformed into an E. coli strain expressing the LacZ beta fragment gene results in <1% white colonies.</p>	Pass
<p>Ligation and Recutting (Terminal Integrity) After a 20-fold over-digestion of Lambda DNA with Aval, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Aval.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Penghua Zhang
Production Scientist
15 Apr 2021



Michael Tonello
Packaging Quality Control Inspector
15 Apr 2021