240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Diluent C
Catalog #: B8003S

 Concentration:
 1X Concentrate

 Lot #:
 0011703

 Assay Date:
 03/2017

 Expiration Date:
 03/2020

 Storage Temp:
 -20°C

Composition (1X): 250 mM NaCl, 10 mM Tris-HCl, 1 mM DTT, 0.1 mM EDTA, 50 % Glycerol, 0.15 % Triton X-100, 200 µg/ml BSA,

(pH 7.4 @ 25°C)

Specification Version: PS-B8003S v1.0 Effective Date: 15 May 2018

Assay Name/Specification (minimum release criteria)	Lot #0011703
Endonuclease Activity (Nicking) - A 50 μ l reaction in CutSmart® Buffer containing 1 μ g of supercoiled PhiX174 DNA and a minimum of 10 μ l of Diluent C incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Non-Specific DNase Activity (16 Hour) - A 50 μ l reaction in CutSmart® Buffer containing 1 μ g of PhiX174-HaeIII DNA and a minimum of 10 μ l of Diluent C incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
pH (buffers/solutions) - The pH of 1X Diluent C is between pH 7.3 and 7.5 at 25°C.	Pass
qPCR DNA Contamination (<i>E. coli</i> Genomic) - A minimum of 1 μ l of Diluent C is screened for the presence of <i>E. coli</i> genomic DNA using SYBR® Green qPCR with primers specific for the <i>E. coli</i> 16S rRNA locus. Results are quantified using a standard curve generated from purified <i>E. coli</i> genomic DNA. The measured level of <i>E. coli</i> genomic DNA contamination is ≤ 1 <i>E. coli</i> genome.	Pass
RNase Activity (Extended Digestion) - A 10 μ l reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 μ l of Diluent C is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	Pass

Authorized by Derek Robinson 15 May 2018







Inspected by
Tony Spear-Alfonso

30 Mar 2017