

www.neb.com info@neb.com



## New England Biolabs Certificate of Analysis

Product Name: NEBNext® Multiplex Oligos for Illumina® (96 Unique Dual Index Primer

Pairs Set 2)

Catalog Number: E6442L
Packaging Lot Number: 10236106
Expiration Date: 03/2026
Storage Temperature: -20°C

Specification Version: PS-E6442S/L v1.0

| NEBNext® Multiplex Oligos for Illumina® (96 Unique Dual Index Primer Pairs Set 2) Component List |                                               |            |                      |
|--------------------------------------------------------------------------------------------------|-----------------------------------------------|------------|----------------------|
| <b>NEB Part Number</b>                                                                           | Component Description                         | Lot Number | Individual QC Result |
| E6612AVIAL                                                                                       | NEBNext® Adaptor for Illumina®                | 10236109   | Pass                 |
| E6610AAVIAL                                                                                      | USER Enzyme                                   | 10236108   | Pass                 |
| E6443AVIAL                                                                                       | NEBNext® Unique Dual Index Primer Set 2 Plate | 10236107   | Pass                 |

| Assay Name/Specification                                                                                                                                                                                                                                                                                                                                                                                                                        | Lot # 10236106 |
|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------|
| * Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Multiplex Oligos for Illumina® (96 Unique Dual Index Primer Pairs Set 2) and meet the designated specifications.                                                                                                                                                                                                       | Pass           |
| Functional Testing (Library Construction, Oligo)  Each of the components are functionally validated through construction of libraries made from commercially available genomic DNA. Libraries are made from each unique barcoded oligonucleotide and are sequenced together on the same Illumina® flow cell and compared across various metrics including library yield, fraction of reads aligning to the reference, GC bias, and insert size. | Pass           |

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



E6442L / Lot: 10236106

Page 1 of 2



Melissa Cormier Production Scientist 25 Apr 2024

issa alia

Michael Tonello

Packaging Quality Control Inspector

24 Jun 2024