

## New England Biolabs Certificate of Analysis


Product Name: NEBNext® Ultra™ II RNA Library Prep Kit for Illumina®  
 Catalog Number: E7770L  
 Packaging Lot Number: 10061672  
 Expiration Date: 07/2021  
 Storage Temperature: -20°C  
 Specification Version: PS-E7770S/L v1.0

NEBNext® Ultra™ II RNA Library Prep Kit for Illumina® Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7764AAVIAL	Nuclease-free Water	10061678	Pass
E7763AAVIAL	0.1X TE	10061685	Pass
E7762AAVIAL	NEBNext® Adaptor Dilution Buffer	10061684	Pass
E7761AAVIAL	NEBNext® First Strand Synthesis Enzyme Mix	10061683	Pass
E7649AAVIAL	NEBNext® Ultra™ II Q5® Master Mix	10061682	Pass
E7648AAVIAL	NEBNext® Ultra™ II Ligation Master Mix	10061681	Pass
E7647AAVIAL	NEBNext® Ultra™ II End Prep Reaction Buffer	10061680	Pass
E7646AAVIAL	NEBNext® Ultra™ II End Prep Enzyme Mix	10061679	Pass
E7427AAVIAL	NEBNext® Second Strand Synthesis Reaction Buffer, 10X	10061677	Pass
E7425AAVIAL	NEBNext® Second Strand Synthesis Enzyme Mix	10061676	Pass
E7422AAVIAL	Random Primers	10061675	Pass
E7421AAVIAL	NEBNext® First Strand Synthesis Reaction Buffer	10061674	Pass
E7374AAVIAL	NEBNext® Ligation Enhancer	10061673	Pass

Assay Name/Specification	Lot # 10061672
<b>Functional Testing (Library Construction, RNA)</b> Each set of reagents is functionally validated and compared to the previous lot through construction of libraries made from commercially available RNA, using the kit's minimum and maximum input requirements. Libraries made from the previous and current lots for both input RNA amounts are sequenced together on the same Illumina flow cell and compared across various metrics including library yield, individual transcript abundance correlations (low vs. high input, old lot vs. new lot), 5'-3' transcript coverage, and fraction of reads mapping to a reference.	Pass
<b>* Individual Product Component Note</b> Standard Quality Control Tests are performed for each component included in NEBNext®	Pass

Assay Name/Specification	Lot # 10061672
Ultra™ II RNA Library Prep Kit for Illumina® and meet the designated specifications.	

This product has been tested and shown to be in compliance with all specifications.



Melissa Arn  
Production Scientist  
17 Jan 2020



Josh Hersey  
Packaging Quality Control Inspector  
23 Mar 2020