

www.neb.com info@neb.com



New England Biolabs Certificate of Analysis

Product Name: DNase I-XT
Catalog Number: M0570L
Concentration: 2,000 U/ml

Unit Definition: One unit is defined the amount of enzyme required to release 260

pmol of FAM from FAM-BHQ1 labeled 35 nt hairpin oligo in 1 min at

30°C in a 50 μl reaction.

Packaging Lot Number: 10245593
Expiration Date: 02/2026
Storage Temperature: -20°C

Storage Conditions: 10 mM Tris-HCl, 2 mM CaCl2, 50% Glycerol, (pH 7.6 @ 25°C)

Specification Version: PS-M0570S/L v2.0

| DNase I-XT Component List | | | | |
|---------------------------|----------------------------|------------|----------------------|--|
| NEB Part Number | Component Description | Lot Number | Individual QC Result | |
| M0570LVIAL | DNase I-XT | 10226674 | Pass | |
| B0570SVIAL | DNase I-XT Reaction Buffer | 10211723 | Pass | |

| Assay Name/Specification | Lot # 10245593 |
|---|----------------|
| Protein Purity Assay (SDS-PAGE) DNase I-XT is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection. | Pass |
| RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 2 units of DNase I-XT is incubated at 37°C. After incubation for 4 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection. | Pass |
| qPCR DNA Contamination (Eukaryotic Genomic) A minimum of 2 units of DNase I-XT is screened for the presence of eukaryotic genomic DNA using SYBR® Green qPCR with universal primers for the 18S rRNA locus. Results are quantified using a standard curve generated from purified E. album genomic DNA. The measured level of eukaryotic genomic DNA contamination is ≤ 2.5 pg DNA/μI. | Pass |

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit



M0570L / Lot: 10245593



www.neb.com/trademarks for additional information.

Heidi Church **Production Scientist**

Heid Chun

26 Feb 2024

Josh Hersey Packaging Quality Control Inspector

18 Jun 2024



M0570L / Lot: 10245593

Page 2 of 2