

www.neb.com info@neb.com



## New England Biolabs Certificate of Analysis

Product Name: Vaccinia Capping System

Catalog Number: M2080S Concentration: 10,000 U/ml

Unit Definition: One unit of Vaccinia Capping Enzyme is defined as the amount of

enzyme required to incorporate 10 pmol of (α<sup>32</sup>P) GTP into an 80 nt

transcript in 1 hour at 37°C.

Packaging Lot Number: 10245746
Expiration Date: 02/2026
Storage Temperature: -20°C

Storage Conditions: 100 mM NaCl , 20 mM Tris-HCl (pH 8.0), 1 mM DTT , 0.1 mM EDTA , 50 %

Glycerol, 0.1 % Triton®X-100

Specification Version: PS-M2080S v1.0

Vaccinia Capping System Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
N2080AVIAL	GTP	10241999	Pass	
M2080SVIAL	Vaccinia Capping System	10239810	Pass	
B9003SVIAL	S-adenosylmethionine (SAM)	10233985	Pass	
B2080AVIAL	10X Capping Buffer	10228128	Pass	

Assay Name/Specification	Lot # 10245746
Endonuclease Activity (Nicking)	Pass
A 50 μl reaction in Capping Buffer containing 1 μg of supercoiled PhiX174 DNA and a	
minimum of 10 units of Vaccinia Capping System incubated for 4 hours at 37°C results	
in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	
Exonuclease Activity (Radioactivity Release)	Pass
A 50 µl reaction in Capping Buffer containing 1 µg of a mixture of single and	
double-stranded [ ³H] E. coli DNA and a minimum of 10 units of Vaccinia Capping	
System incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	
Protein Purity Assay (SDS-PAGE)	Pass
Vaccinia Capping System is ≥ 95% pure as determined by SDS-PAGE analysis using	
Coomassie Blue detection.	
RNase Activity (Extended Digestion)	Pass
A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA	



M2080S / Lot: 10245746

Page 1 of 2



Assay Name/Specification	Lot # 10245746
and a minimum of 10 units of Vaccinia Capping System is incubated at 37°C. After incubation for 4 hours, >90% of the substrate RNA remains intact as determined by	
gel electrophoresis using fluorescent detection.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Jessica Cane Production Scientist

14 Feb 2024

Josh Hersey

Packaging Quality Control Inspector

05 Jun 2024

