

www.neb.com info@neb.com



## New England Biolabs Certificate of Analysis

Product Name: Kasl
Catalog Number: R0544S
Concentration: 5,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of pBR322 DNA in rCutSmart Buffer in 1 hour at 37°C in a total

reaction volume of 50 μl.

Packaging Lot Number: 10239975
Expiration Date: 04/2025
Storage Temperature: -80°C

Storage Conditions: 10 mM Tris-HCl, 100 mM KCl, 1 mM DTT, 0.1 mM EDTA, 200 µg/ml

rAlbumin, 50% Glycerol, (pH 7.4 @ 25°C)

Specification Version: PS-R0544S/L v4.0

Kasl Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
R0544SVIAL	Kasl	10239974	Pass	
B6004SVIAL	rCutSmart™ Buffer	10238051	Pass	

Assay Name/Specification	Lot # 10239975
Blue-White Screening (Terminal Integrity) A sample of LITMUS38i vector linearized with a 10-fold excess of KasI, religated and transformed into an E. coli strain expressing the LacZ beta fragment gene results in <1% white colonies.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in rCutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [ ³H] E. coli DNA and a minimum of 5 units of Kasl incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
<b>Ligation and Recutting (Terminal Integrity)</b> After a 20-fold over-digestion of pBR322 DNA with Kasl, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Kasl.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in rCutSmart™ Buffer containing 1 µg of pBR322 DNA and a minimum of 5 units of Kasl incubated for 16 hours at 37°C results in a DNA pattern free of	Pass



R0544S / Lot: 10239975

Page 1 of 2



Assay Name/Specification	Lot # 10239975
detectable nuclease degradation as determined by agarose gel electrophoresis.	
Protein Purity Assay (SDS-PAGE) Kasl is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	Pass
qPCR DNA Contamination (E. coli Genomic)  A minimum of 5 units of Kasl is screened for the presence of E. coli genomic DNA using SYBR® Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Ana Egana **Production Scientist** 

03 Jun 2024

Michael Tonello

Packaging Quality Control Inspector

03 Jun 2024



R0544S / Lot: 10239975

Page 2 of 2