


## New England Biolabs Certificate of Analysis

Product Name: Monarch Plasmid Miniprep Columns  
 Catalog Number: T1017L  
 Concentration: 1 X Concentrate  
 Lot Number: 10029801  
 Expiration Date: 04/2021  
 Storage Temperature: 25°C  
 Specification Version: PS-T1017L v1.0

Assay Name/Specification	Lot # 10029801
<b>Functional Testing (DNA Recovery and Purity)</b> Twenty-four Monarch® Plasmid Miniprep Columns are tested with 1.5 µg of 1 Kb Ladder resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.



Ana Egana  
Production Scientist  
07 May 2019



Heidi Church  
Packaging Quality Control Inspector  
07 May 2019