


New England Biolabs Certificate of Analysis

Product Name: Monarch gDNA Removal Columns
Catalog Number: T2017L
Packaging Lot Number: 10044096
Expiration Date: 06/2021
Storage Temperature: 25°C
Specification Version: PS-T2017L v1.0

Assay Name/Specification	Lot # 10044096
Functional Testing (DNA Recovery) Eight Monarch® gDNA Removal Columns are tested with 10 µg of Lambda/BstEII Marker resulting in ≥65% recovery in ≥80% of the samples.	Pass

This product has been tested and shown to be in compliance with all specifications.



Ana Egana
 Production Scientist
 19 Dec 2019



Heidi Church
 Packaging Quality Control Inspector
 19 Dec 2019