

New England Biolabs Certificate of Analysis


Product Name: *Monarch® RNA Cleanup Kit (50 µg)*
 Catalog Number: *T2040S*
 Packaging Lot Number: *10144723*
 Expiration Date: *09/2024*
 Storage Temperature: *25°C*
 Specification Version: *PS-T2040S/L v3.0*

Monarch® RNA Cleanup Kit (50 µg) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T2047-21	RNA Cleanup 10 Columns (50 µg)	10137052	Pass
T2042-21	Monarch® RNA Cleanup Wash Buffer	10144952	Pass
T2041-21	Monarch® RNA Cleanup Binding Buffer	10144940	Pass
T2018-21	Monarch® Collection Tubes II	10041732	Pass
B1500-21	Nuclease-free Water	10144950	Pass

Assay Name/Specification	Lot # 10144723
<p>Functional Testing (Enzyme Inactivation and Removal) Eight Monarch® RNA Cleanup Columns (T2047, 50 µg) are tested with 5 µg of RNA and 2 units of DNase I. A 50 µl reaction in DNase I Reaction Buffer with 1 µg of pBR322 and 50% of the column eluate incubated for 10 minutes at 37°C results in >99.95% removal of DNase I activity in ≥75% of the samples.</p>	Pass
<p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in Monarch® RNA Cleanup Kit (50 µg) and meet the designated specifications.</p>	Pass
<p>Functional Testing (RNA Recovery and Purity) Twenty-four Monarch® RNA Cleanup Columns (T2047, 50 µg) are tested with 5 µg of RNA resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Ana Egana
Production Scientist
28 Oct 2022



Michael Burzyk
Packaging Quality Control Inspector
28 Oct 2022