

New England Biolabs Certificate of Analysis

Product Name: Monarch® gDNA Purification Columns

Catalog Number: T3017L 10244982 Packaging Lot Number: Expiration Date: 03/2026 Storage Temperature: 25°C

PS-T3017S/L v2.0 Specification Version:

Monarch® gDNA Purification Columns Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
T3017-1	gDNA Purification 50 Columns	10228338	Pass	
T2018-1	Monarch® Collection Tubes II	10183685	Pass	

Assay Name/Specification	Lot # 10244982
Functional Testing (DNA Recovery and Purity)	Pass
Twenty-four Monarch® gDNA Purification Columns are tested with 5 µg of HindIII	
digested Lambda resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and	
260/230 are ≥1.75 in ≥80% of the samples.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Darcie Spaulding Production Scientist 29 May 2024

kaging Quality Control Inspector

29 May 2024



T3017L / Lot: 10244982