

320 Newburyport Turnpike Rowley, MA 01969-2002 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name:	Monarch® gDNA Purification Columns
Catalog Number:	T3017L
Packaging Lot Number:	10244983
Expiration Date:	03/2026
Storage Temperature:	25°C
Specification Version:	PS-T3017S/L v2.0

Monarch® gDNA Purification Columns Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
T3017-1	gDNA Purification 50 Columns	10228338	Pass	
T2018-1	Monarch® Collection Tubes II	10146188	Pass	

Assay Name/Specification	Lot # 10244983
Functional Testing (DNA Recovery and Purity)	Pass
Twenty-four Monarch® gDNA Purification Columns are tested with 5 µg of HindIII	
digested Lambda resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and	
260/230 are ≥1.75 in ≥80% of the samples.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Darri Apaulding

Darcie Spaulding Production Scientist 30 May 2024

June Gaumont Packaging Quality Control Inspector 30 May 2024



T3017L / Lot: 10244983 Page 1 of 1